

Minutes of the meeting of the Pre-Bid Conference (PBC) held on 30.09.2026 for the proposed procurement of "Infrasound Sensor"

Members present

1. Dr. Abhey Ram Bansal (Chairman)
2. Shri.Vijaya Raghavan (Member)
3. Dr.Monalisa Mallick (Member)
4. Venkatesh Vempati(IO)
5. Dr. Rajesh Rekapalli (Co-PI & IO(AWS))

S&P team provided necessary assistance.

The representatives of the following firms attended the PBC-

1. M/s Spry Technocon Private Limited, Gurugram, Haryana, India (Online)
2. M/s Chrisvin Geomet Services Pvt. Ltd., 826, Ground Floor, Tarapore Towers, Anna Salai, Chennai – 600002
3. M/s Aimil Ltd, S D Road, Secunderabad.
4. National Commodities Management Services Limited, Nagole, Hyderabad.

The following points have been discussed during the PBC for Infra sound sensor/microbarometer and the response of the CSIR-NGRI to the queries raised by the bidders are as under

Sl. No.	Point No. of tender Document	Query raised by the Bidder	Response of CSIR-NGRI
M/s Spry Technocon private limited, Gurugram, Haryana India			
1.	Chapter 4, Clause 4.2 - Specification, Sl. No. 4,	Requested Amendment: Kindly amend the self-noise requirement in tender: "Self-Noise at 1 Hz shall not exceed the GLW (Global Low Wind) curve by more than 20 dB."	This is the minimum basic criteria required for the project. As we use the windshield for the infrasound sensors, wind noise will remain low. No amendment required.

The bidders informed that they did not have any other point/query to be discussed and PBC got concluded.

For IO & PL

Monalisa Mallick

(Dr. Rajesh Rekapalli)

Rsu

CSIR-NGRI team informed that the minutes of the PBC, along with clarifications and changes/amendments, if any, agreed by CSIR-NGRI shall be uploaded on CPPP/GeM and CSIR-NGRI website shortly for information and reference of all the participating bidders. All the bidders are requested to take a note of the clarifications/ changes agreed by CSIR-NGRI after the PBC, which shall be treated as an amendment to the existing tendered specifications and related requirements.

IO& Convenor (Mr. V. Venkatesh)	PL (Dr. Sandeep Gupta)	Chairperson (Dr Abhey Ram Bansal)
<i>For</i> 	<i>For</i> 	
Member (Shri. R. Vijayaraghavan)	Member (Dr. Monalisa Mallick)	IO& Convenor (AWS) (Mr.R.Rajesh)
	<i>Monalisa Mallick</i>	

Amendment to the existing specifications and related requirements for the proposed procurement of "Broadband Seismographs and Infrasound Sensor/Microbarometer" subsequent to the Pre-Bid Conference (PBC) held on 30.09.2026.

Sl. No.	Point No. of tender Document	Existing specifications	Amended specifications
No Amendments			

IO& Convenor (Mr. V. Venkatesh)	PL (Dr. Sandeep Gupta)	Chairperson (Dr Abhey Ram Bansal)
		
Member (Shri. R. Vijayaraghavan)	Member (Dr. Monalisa Mallick)	
		

